



RE: Hach Company digital ORP probe warranty questionnaire

We are sorry to hear that you are having issues with your GLI differential ORP probe. The sensor warranty covers the instrument against defects in materials and craftsmanship for 12 months, but also includes a pro-rated warranty that covers the sensor for 30 months from the instrument sale date for other failures.

If you would like to apply for the sensor warranty or pro-rate, please fill in the information below and email the completed form to the technical support team at techhelp@hach.com.

Section 1: Customer information	
Name:	
Phone number:	
Company:	
Hach account number:	
Shipping address:	
Billing address:	
Section 2: Probe information	
Probe part number:	
Probe serial number/order number:	/
Reading in your process:	
Expected value:	
Brief description of issues:	
Section 3: Clean the sensor	
1. Clean the sensor by removing loose debris with a stream of water and soft bristled brush	
2. Prepare a mild soap solution with warm water and lanolin free soap	
3. Soak the sensor for 2-3 minutes in the solution	
4. Use a small soft bristled brush and clean the electrode and salt bridge	
5. If deposits remain after cleaning with soap, use a dilute acid, such as muriatic acid to soak the probe for 5 minutes	
6. Soak the sensor for 2-3 minutes in the soap solution	
7. Rinse sensor with clean, warm water	

Section 4: Sensor calibration	
Perform a 1-point sensor calibration and record readings from controller	
200 mV standard	<input type="text"/>
Section 5: Replace salt bridge and standard cell solution	
Following the salt bridge and standard cell solution replacement, perform a 1-point sensor calibration and record the readings from the controller	
200 mV standard	<input type="text"/>
Section 6: Troubleshooting	
1. Place the sensor in a fresh 200 mV standard. Allow for the temperature to equilibrate	
2. On the controller, enter the 'Sensor Setup' menu, highlight 'Diag/Test', press enter	
3. Highlight 'Sensor Signal' and press enter. Record value:	<input type="text"/>

Once we receive your completed questionnaire, we will evaluate the information. If deemed a defect in materials or craftsmanship within the warranty period a new sensor will be sent at no cost. If not related to a manufacturing, but within the valid pro-rate period, you will receive a new sensor at a pro-rated price. If you need a new PO for this order, please indicate that below and one of our order specialists will contact you for that PO.

- I will need to give a new PO for the purchase price of this probe
- I will NOT need a new PO for the purchase of this probe

If you have questions, please contact the technical support team via email at techhelp@hach.com or via phone at (800)227-4224 and follow the prompts to speak to a technical support representative.

Thank you for your business!

For internal use only:

Technical advisor name:	<input type="text"/>	Date:	<input type="text"/>
Sensor offset:	<input type="text"/>		<input type="text"/>
Pro-rated replacement (y,n,n/a):	<input type="text"/>	Price:	<input type="text"/>
Warranty replacement (y,n,n/a):	<input type="text"/>		<input type="text"/>
Notes:	<input type="text"/>		